



TECNOSOFT SRL
Via Galvani, 4, 20068, Peschiera Borromeo (MI), ITALY
Tel +39 0226922888 - Fax +39 0226922875
P.IVA e CF IT09553010159
Cap. Soc. € 10.400,00 i.v.
R.I. Milano 291533 - R.E.A. Milano 1305624
N. Registro A.E.E. IT08020000000687
N. Registro Pile IT09060P000000026
email hardware@tecnosoft.eu - web www.tecnosoft.eu



Certificato di calibrazione / Calibration certificate

Strumento / Instrument: Data logger di temperatura / Temperature data logger
Costruttore / Manufacturer: Tecnosoft srl
Modello / Model: SterilDisk
Numero di serie / Serial number: 1001006420
Numero certificato / Certificate number: 6420-16G22

Temperatura di riferimento Reference temperature °C	Temperatura calibrata Calibrated temperature °C	Scostamento dal riferimento Error °C
24,73	24,73	0,00
49,83	49,83	0,00
74,96	74,96	0,00
98,99	98,99	0,00
125,01	125,02	0,01
140,01	140,01	0,00

Data di calibrazione / Calibration date: 22/07/2016

Data consigliata prossima calibrazione / Suggested date for next calibration: 22/07/2017

Strumento di calibrazione / Calibration instrument: Wika Thermostatic Bath CTB9430 S/N:111152037

Termometro campione / Reference thermometer: Platinum Thermo Resistance, connected to:

Thermometer Delta Ohm HD2107.2 s/n 10008622 with calibration certificate LAT 169 0457/16 of 18/03/2016