

Allgemein / General / Général / General:

Gerät / Module type / Modèle / Modelo:	testo 184 T1
Serien-Nr. / Serial no. / No. de série / Número de serie:	43810000
Messbereich / Measuring range / Etendue de mesure / Rango de medición :	-35.0 ... +70.0 °C
Zulässige Toleranz / Permissible tolerance / Tolérance admise / Tolerancia permitida:	± 0.5 °C

Sollwert / Reference / Référence / Referencia:

Type	Certificate	Uncertainty(k=2)
Agilent34410A	34410AMY53003319	0.1 Ohm

Sensor / Sensor / Capteur / Capteur

Type	Certificate	Accuracy @5.0°C
TT3-5KCs-80	4711Z 08/14	± 0.1 °C

Elektronische Kalibrierung / Electronic calibration / Etalonnage électronique / Calibración electrónica

No.	Reference(°C)	Reading(°C)	Deviation(°C)
1	-30.02	-29.94	0.08
2	-10.02	-10.04	-0.02
3	5.01	5.00	-0.01
4	24.98	24.97	-0.01
5	69.97	69.99	0.01

Systemunsicherheit / System uncertainty / L'incertitude du système / Incertidumbre system

Messwerte / Measured values / Valeurs mesurées / Valores medidos:		
Sollwert / Reference / Référence / Referencia:	Zulässige Toleranz / Permissible tolerance / Tolérance admise / Tolerancia permitida:	Unsicherheit / Uncertainty / Incertitude / Incertidumbre:
5.0 °C	± 0.5 °C	0.06 °C