

Certificate Data Logger testo 184



Allgemeine / General / Général / General:

Gerät / Module type / Modèle / Modelo:	testo 184 T3
Serien-Nr. / Serial no. / No. de série / Número de serie:	44000128
Meßbereich / Measuring range / Etendue de mesure / Rango de medición :	-35.0 ... +70.0 °C
Zulässige Toleranz / Permissible tolerance / Tolérance admise / Tolerancia permitida:	± 0.5 °C

Sollwert / Reference / Référence / Referencia:

Type	Certificate	Uncertainty(k=2)
Agilent34410A	34410AMY53003319	0.1 Ohm

Sensor / Sensor / Capteur / Capteur

Type	Certificate	Accuracy @5.0°C
TT3-5KCs-80	4400Z 11/13	± 0.1 °C

Elektronische kalibrierung / Electronic calibration / Etalonnage électronique / Calibración electrónica

No.	Reference(°C)	Reading(°C)	Deviation(°C)
1	-30.02	-30.11	-0.09
2	-10.02	-9.98	0.04
3	4.98	5.02	0.04
4	24.99	25.02	0.03
5	69.98	70.07	0.09

System unsicherheit / System uncertainty / L'incertitude du système / Incertidumbre system

Meßwerte / Measured values / Valeurs mesurées / Valores medidos:		
Sollwert / Reference / Référence / Referencia:	Zulässige toleranz / Permissible tolerance / Tolérance admise / Tolerancia permitida:	Unsicherheit / Uncertainty / Incertitude / Incertidumbre:
5.0 °C	± 0.5 °C	0.06 °C