

# Certificate Data Logger testo 184



## Allgemeine / General / Général / General:

Gerät / Module type / Modèle / Modelo:	testo 184 T4
Serien-Nr. / Serial no. / No. de série / Número de serie:	44100644
Meßbereich / Measuring range / Etendue de mesure / Rango de medición :	-80.0 ... +70.0 °C
Zulässige Toleranz / Permissible tolerance / Tolérance admise / Tolerancia permitida:	± 0.8 °C ( -80.0 ... -35.1 °C) ± 0.5 °C ( -35.0 ... +70.0 °C)

## Sollwert / Reference / Référence / Referencia:

Type	Certificate	Uncertainty(k=2)
Agilent34410A	34410AMY53003319	0.1 Ohm

## Sensor / Sensor / Capteur / Capteur

Type	Certificate	Accuracy @5.0°C
PT1000	1/3 Class B	± 0.1 °C

## Elektronische kalibrierung / Electronic calibration / Etalonnage électronique / Calibración electrónica

No.	Reference(°C)	Reading(°C)	Deviation(°C)
1	-69.87	-69.90	-0.03
2	-9.99	-10.11	-0.12
3	5.05	5.01	-0.04
4	25.09	25.03	-0.06
5	70.10	70.06	-0.04

## System unsicherheit / System uncertainty / L'incertitude du système / Incertidumbre system

Meßwerte / Measured values / Valeurs mesurées / Valores medidos:		
Sollwert / Reference / Référence / Referencia:	Zulässige toleranz / Permissible tolerance / Tolérance admise / Tolerancia permitida:	Unsicherheit / Uncertainty / Incertitude / Incertidumbre:
5.0 °C	± 0.5 °C	0.08 °C